

ORDER INFORMATION

CODE : 1x15ml, 1x5ml (20ml)
1x30ml, 1x10ml (40ml)

HDL - CHOLESTEROL

DIRECT ENZYMATIC CHOLORIMETRIC

INTENDED USE :

This reagent kit is intended for "*in vitro*" quantitative determination of High - Density Lipoprotein Cholesterol (HDL-C) in serum & plasma.

CLINICAL SIGNIFICANCE :

HDL particles serve to transport in the blood-stream. HDL is known as "good cholesterol" because high levels are thought to lower the risk of heart disease and coronary artery disease. A low HDL cholesterol level, is considered a greater heart disease risk.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

PRINCIPLE :

Direct determination of serum HDL (high-density lipoprotein cholesterol) levels without the need for any pre-treatment or centrifugation of the sample.

The method depends on the properties of a detergent which solubilizes only the HDL so that HDL-c is released to react with the cholesterol esterase, cholesterol oxidase and chromogens to give colour. The non HDL lipoprotein LDL, VLDL and chylomicrons are inhibited from reacting with the enzymes due to absorption of the detergents on their surfaces. The intensity of the color formed is proportional to the HDL concentration in the sample.

REAGENT COMPOSITION :

Reagent 1 : Buffer Reagent
Reagent 2 : Substrate Reagent
HDL-C Calibrator : Concentration Printed on the Vial.

SAMPLES :

Serum or heparinized plasma, free of hemolysis, removed from the blood clot as soon as possible.
Anticoagulants containing citrate should not be used.

WORKING REAGENT PREPARATION & STABILITY :

Reagent 1 & Reagent 2 are Ready to use and are stable up to the expiry date stated on the label.
HDL-C Calibrator : Dissolve with distilled water (Qty Printed on the vial). Cap and mix gently to dissolve contents. Reconstituted calibrator is stable for 7 days at 2° to 8°C or 21 days at -20°.

GENERAL SYSTEM PARAMETERS :

Reaction type	Endpoint
Wave length	546 nm
Light Path	1 Cm
Reaction Temperature	37°C
Blank / Zero Setting	Reagent
Reagent Volume	1 ml
Sample Volume	10 µl
Incubation Time	10 Min.
Calibrator Concentration	Printed on vial
Low Normal at 37°C	35 mg/dl
High Normal at 37°C	88 mg/dl
Linearity	200 mg/dl

ASSAY PROCEDURE :

	Blank	Calibrator	Sample
Reagent 1	750 µl	750 µl	750 µl
Calibrator		10 µl	
Sample			10 µl

Mix and incubate for 5 Minutes at 37°C and then add

Reagent 2	250 µl	250 µl	250 µl
-----------	--------	--------	--------

Mix and read the optical density (A) after 10 minutes incubation at 37°C.

CALCULATION :

$$\text{HDL-C (mg/dl)} = \frac{\text{OD of Sample}}{\text{OD of Calibrator}} \times \text{Conc. of Calibrator}$$

LINEARITY :

Reagent is Linear up to 200 mg/dl.
Dilute the sample appropriately and re-assay if HDL-C Concentration exceeds 200 mg/dl.

REFERENCE NORMAL VALUE :

Adult male: 35 -79.5 mg/dl
Adult female: 42 -88 mg/dl
It is recommended that each laboratory should assign its own normal range.

QUALITY CONTROL :

For accuracy it is necessary to run known controls with every assay.

LIMITATION & PRECAUTIONS :

1. Storage conditions as mentioned on the kit to be adhered.
2. Do not freeze or expose the reagents to higher temperature as it may affect the performance of the kit.
3. Before the assay bring all the reagents to room temperature.
4. Avoid contamination of the reagent during assay process.
5. Use clean glassware free from dust or debris.

BIBLIOGRAPHY :

1. Natio H K Cholesterol Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984; 1207-1213 and 437.
2. US National Cholesterol Education Program of the National Institutes of Health.
3. Young DS. Effects of Drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
4. Young DS. Effects of diseases on Clinical Lab. Tests 4th ed AACC 2001.
5. Burlis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
6. Tietz N W et al, Clinical to Laboratory Tests, 3rd ed AACC 1995.



ASHSKY CHEM INDIA PVT LTD